



16F Arc Place Bldg.  
142, Teheran-ro, Gangnam-gu,  
Seoul, 06236, Korea  
**bd.com**

---

**Title : Suspension of sale and Recall - BD Emerald Syringe 302923, Lot No. 8344699)**

**Date : August 27, 2019**

**To : Attention of Final User Facility**

**From : BD Korea**

---

1. We wish you continued development and great prosperity.
2. According to Article 31-5, Article 34 of the Medical Devices Act, the following medical device should be applied to “Suspension of sale and Recall”.

Product Name	MFDS License	Cat. No.	Lot No.	Q'ty	Description of the problem
BD Emerald Syringe	Suhuh 12-1761	302923	8344699	1,200	Foreign substances(caterpillars) generated during Korea domestic distribution

3. **BD Korea** is conducting a medical device recall for Lot No. 8344699 of BD Emerald Syringe due to foreign substances(caterpillars) generated during domestic distribution. While this would likely be identified prior to use, BD is to inform you that this is for better and safer use.

**4. Please Take the Following Actions:**

- 1) Immediately review your inventory for the specific catalog and lot numbers listed above.
- 2) Please return product to **BD Korea**.
- 3) Complete the attached Customer Response Form and return to the **BD Korea** contact noted on the form.
- 4) If you have any question, please contact to **BD Korea**. (**BD Korea ☎ 02-3404-3700**)

5. **BD** is committed to advancing the world of health. Our primary objectives are patient and user safety and providing you with quality products. We apologize for any inconvenience this issue may have caused you and thank you in advance for helping us to resolve this matter as quickly and effectively as possible.

**BD Korea**