

Standard Measures

NOTE: Reporting Person refers to the manufacturer, importer, supplier or registrant who submitted the FSCA Notification Report to HSA.

Pursuant to the HPA, Reporting Person shall ensure all the following measures are undertaken in relation to this FSCA:

- a. Subject to HSA's concurrence, medical devices that are affected by the FSCA shall be determined from the product owner's Field Safety Notice (FSN);
- b. New supply of correction-in-progress medical devices may proceed, subject to the following conditions:
 - A copy of the Field Safety Notice (FSN) and/or, where required, a Dear Healthcare Professional Letter (DHCPL) shall be provided to each local consignee, prior to supply;
 - ii) Once the product owner's corrective and preventive actions (CAPA) has been finalized and made available for implementation in any country, the CAPA shall be submitted to the HSA for review immediately. Upon CAPA availability, all correction-in-progress devices shall undergo full correction without further delay;
 - iii) Full and effective compliance and performance of all measures instructed by the Authority as part of this Notice or future Notices in relation to this FSCA; and
 - iv) Prior to closure of this FSCA, Reporting Person shall submit a declaration on letter head to confirm that full correction for all correction-in-progress devices supplied in Singapore has been completed;
- c. This Notice is issued without prejudice to any precedent or subsequent notices issued by the Authority instructing a cease to the manufacture, import, supply or use of the affected medical device;
- d. Supply of medical devices not affected by this FSCA or supply of corrected medical devices may proceed subject to any other applicable regulatory requirements in Singapore being complied with;
- e. Reporting Person may supply by export affected devices out of Singapore, subject to marketing authorisation. Reporting Person should notify the appropriate regulatory authority of the country of import regarding the exportation of devices affected by the FSCA;
- f. Notify all consignees with existing supply of affected medical devices immediately of this FSCA. For the purposes of this Notice, 'existing supply' refers to any affected MDs supplied to consignees in Singapore prior to Date of Notification of FSCA to HSA. Medical devices imported but not supplied prior to Date of Notification of FSCA to HSA shall not be considered to be part of existing supply. Any supply, other than existing supply, of affected medical devices will hereinafter be referred to as 'new supply';
- g. Written acknowledgements of receipt of the product owner's FSN shall be obtained from consignees supplied with correction-in-progress medical devices. Records of acknowledgement receipts shall be provided to HSA when requested;
- h. Where on-site device modification is required, written acknowledgement (e.g. service report) confirming effective completion of the field corrections required as per product owner's FSN shall be obtained from consignees supplied with correction-in-progress medical devices. Written records of acknowledgements shall be provided to HSA when requested;
- For implantable medical devices, written acknowledgement of receipt of the product owner's Dear Healthcare Professional Letter (DHCPL) shall be obtained from qualified practitioners who have implanted the affected medical devices. Records of acknowledgement receipts shall be provided to HSA when requested;

- j. Where notified by HSA that the DHCPL or FSN requires amendment, Reporting Person shall ensure that the necessary amendments are made, and proceed to disseminate the amended DHCPL or FSN only upon receipt of HSA's written permission to do so;
- k. Ensure that documents requested in relation to this FSCA are submitted within the requested timelines in HSA's Notices. If necessary, measures (other than those in the product owner's FSN) may be imposed on corrected devices. A failure to provide supporting documents for CAPA effectiveness, inter alia, may lead to imposition of additional measures on corrected devices;
- I. (Applicable to registered devices only) Prior to new supply of corrected medical devices, verify whether a Change Notification (CN) submission through the Medical Device Information & Communication System (MEDICS) is required. Refer to GN-21: Guidance on Change Notification for further information. If a MEDICS CN submission and approval is necessary, new supply shall not proceed unless prior approval from HSA has been received. Any new supply shall be subject to the conditions or restrictions that may be imposed as part of HSA's approval. Measures to ensure strict compliance to these conditions or restrictions shall be undertaken by Reporting Person. More information on CN is available on our website at https://www.hsa.gov.sg/.

NOTE: For avoidance of doubt, correction-in-progress device is a medical device which is affected by the FSCA and has not completed the CAPA intended by the product owner to wholly correct for the FSCA. For example, a medical device affected by a FSCA and only having interim risk mitigation measures implemented will be considered as a correction-in-progress device.

NOTE: Subject to HSA's concurrence, medical devices that have completed the full correction required by the product owner's FSN shall be deemed corrected devices. Supply of corrected devices may proceed subject to any other applicable regulatory requirements (e.g. change notification) in Singapore that must be complied with. HSA reserves its rights to impose subsequent measures on corrected devices if deemed necessary.

NOTE: Marketing authorisations would include (i) registration on Singapore Medical Device Register (SMDR), (ii) placement on the Transition List, or (iii) Special Access Route (SAR) for local supply (GN26, GN-27, GN-29 or GN30(CR)).

Please be informed that the penalties that apply for failure to perform the measures required under this Notice are a fine not exceeding \$10,000 or to imprisonment for a term not exceeding 6 months, or to both.